

10780111 • MAN031 09 - 2004 2ªEd.

**IME** 

ISTRUMENTI MISURE ELETTRICHE SpA

Via Travaglia 7  
20094 CORSICO (MI)  
ITALIA

Tel. 02 44 878.1  
Fax 02 45 03 448  
+39 02 45 86 76 63

[www.imeitaly.com](http://www.imeitaly.com)  
[info@imeitaly.com](mailto:info@imeitaly.com)

# DG4Y

°C

°F

**J**

**J**

**K**

**K**

**T**


**T**

**S**

**S**

Uscita dal campo di misura (Vis. fissa)  
Interruzione termoelemento (Vis. lampeggiante)

En dehors de l'étendue de mesure (affichage fixe)  
Rupture du thermoélément (affichage clignotant)



Display overrange / underrange (fixed indication)  
Burn-out (blinking indication)

Meßbereich außerhalb (Anzeige Fest)  
Thermoelementbruch (Anzeige Blinkend)

Uscita campo compensazione giunto freddo

Compensation de la soudure froide en dehors des limites



Cold junction compensation, overrange / underrange

Vergleichsstellentemperatur außerhalb

